



Bishop Meadow Road,
Loughborough,
Leicestershire,
LE11 5RG
Tel: 01509 231166
Fax: 01509 231893

Specification Sheet

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2008 by SAI Global Certificate Number QMS42099

Catalogue Number:	J/2880	Expiry Phrase:	Use within 2 yrs of opening
--------------------------	--------	-----------------------	-----------------------------

Description:	Buffer solution pH10 (Borate)		
Grade:	Buffer Solution	Application:	

Result Name	Specification	Units
pH @ 20°C	≥ 9.98 and ≤ 10.02	pH

Additional Information:	The pH value is determined from the response of a calibrated glass electrode. The pH scale is that defined in BS1647:1984. The glass electrode is calibrated by measuring its response when immersed in reference buffer solutions. Reference buffer solutions are prepared according to the formulations in the British Standard and are traceable to NIST reference material
--------------------------------	--



Mrs K S Cluskey BSc
Quality & Compliance Manager.

Thursday 11th of January 2018

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above. If there are any questions regarding this certificate, please call Technical Support on +32 1456 5600