

Bishop Meadow Road, Loughborough, Leicestershire, LE11 5RG Tel: 01509 231166 Fax: 01509 231893

## **Specification Sheet**

Thermo Fisher Scientific's Quality System has been found to conform to Quality Management System Standard ISO9001:2008 by SAI Global Certificate Number QMS42099

Catalogue Number: J/7620	Expiry Phrase:	Use within 2 yrs of opening
-----------------------------	----------------	-----------------------------

Description:	Sodium hydroxide,M (N) (free from carbon	nate)	
Grade:	SVS	Application:	Standard solution for volumetric analysis

Result Name	Specification	Units
Factor	>= 0.999 and <= 1.001	

Additional	This product is traceable to NIST. Meets the current analytical specification of BP and Ph. Eur. This
Information:	product is tested internally to confirm it is free from carbonate.





Mrs K S Cluskey BSc Quality & Compliance Manager. Thursday 11th of January 2018

Note: The data listed is valid for all package sizes of this lot of this product, expressed as an extension of the catalogue number listed above. If there are any questions regarding this certificate, please call Technical Support on +32 1456 5600